

COMBS SCHOOL OZONE 2010 Data
8 Hour Average Daily Maximum (PPB)

	January	February	March	April	May	June	July	August	September	October	November	December
1				49	43	46	60	38	51	46		
2				53	54	48	54	46	55	53		
3				55	55	44	47	48	50	47		
4				52	59	45	41	50	47	38		
5				48	56	46	42	49	49	37		
6				52	46	42	52	43	39	38		
7				55	54	40	54	49	42	37		
8				59	43	42	47	43	34	45		
9				58	57	39	50	40	38	52		
10				50	58	33	59	44	42	50		
11				49	59	53	43	41	45	47		
12				44	47	55	44	31	52	55		
13				50	53	56	42	50	53	50		
14				53	58	59	41	48	49	47		
15				54	64	59	59	47	45	47		
16				51	62	55	54	53	46	43		
17				57	51	57	52	56	47	45		
18				57	43	54	56	45	49	39		
19				50	58	49	55	54	44	32		
20				44	54	52	48	56	42	30		
21				44	56	47	53	49	39	40		
22				48	54	AE	48	58	33	39		
23				44	51	AE	44	60	40	43		
24				53	51	AE	55	57	45	42		
25				56	56	54	58	48	51	30		
26				56	62	48	49	45	49	35		
27				59	62	46	49	51	55	38		
28				53	42	AE	42	46	49	36		
29				56	63	67	35	34	56	39		
30				45	58	62	47	37	62	31		
31				50			36	34		36		
	# of 8 hr exceedances		0	# of 8 hr exceedances		0	# of 8 hr exceedances		0	# of 8 hr exceedances		0

COMBS SCHOOL OZONE 2010 Data Reference Page

Corresponding Color Codings

AQS Codings

8 hour exceedances	any sample > 75 PPB	
General maintenance	monitor is down for general maintenance activities or service for more than 3 hours	BA
Power failures	power outage for more than 6 hours	AV
CAL / VER / AUDIT	monitor is down to perform monthly verification or annual system calibrations or QC Audit (ADEQ) for more than 3 hours	AT AT AZ
Shelter Temp	Shelter Temperature out of limits	AE
NOT OPERATED	This instrument is only operated during the ozone season April 1 - October 31	

Maximum Value	67
----------------------	----